




| | | | |
|--|--|---------------------------|--|
| Chemsafety Ltd | | Client Number 9360 | |
| PO Box 8141, Riccarton, Christchurch, 8440 2B, 303 Blenheim Road, Upper Riccarton, Christchurch, 8041 Telephone 03 366-3700 www.chemsafety.co.nz | | | |
| Authorised Representative Mrs Livia Fioravanti Laboratory Manager | | | |
| Programme Chemical Testing Laboratory Accreditation Number 1328 Initial Accreditation Date 17 June 2019 | | | |
| Conformance Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories | | | |
| Laboratory Services Summary 2.64 Asbestos | | | |
| Key Technical Personnel Mr Tony Bird 2.64(b)(c)(d) Mrs Livia Fioravanti 2.64(a)(b)(c)(d) | | | |

| | | | |
|--|---------|----------------|-------------|
| Operations Manager Authorisation:  | Issue 8 | Date: 15/11/22 | Page 1 of 2 |
|--|---------|----------------|-------------|



Chemsafety Ltd
Chemical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation Number 1328

2.64 Asbestos

(a) Identification

Bulk identification of asbestos using polarised light microscopy

AS 4964:2004 Method for the Qualitative Identification of Asbestos in Bulk Samples

(b) Fibre Counts

Determination of airborne asbestos fibre counts on membrane filters

NOHSC:3003: National Occupational Health and Safety Commission Guidance Note on the Membrane Filter Method for Estimating Airborne Asbestos 2nd Edition [NOHSC:3003 (2005)]

(c) Field Sites

Establishment and operation of field sites for the purposes of asbestos fibre counting, in accordance with procedures listed in 2.64(b)

(d) Field Testing

Performance of field testing for the purposes of asbestos fibre counting, in accordance with procedures listed in 2.64(b)

| | | | | |
|--------------------------------------|--|---------|---------------|-------------|
| Operations Manager Authorisation: | | Issue 8 | Date:15/11/22 | Page 2 of 2 |
|--------------------------------------|--|---------|---------------|-------------|