



**Chemsafety Ltd**

**Client Number 9360**

PO Box 8141, Riccarton, Christchurch, 8440  
2B, 303 Blenheim Road, Upper Riccarton, Christchurch, 8041

**Telephone 03 366-3700**

**www.chemsafety.co.nz**

**Authorised Representative**

Mrs Livia Fioravanti  
Laboratory Manager

**Programme**

Chemical Testing Laboratory

**Accreditation Number 1328**

**Initial Accreditation Date 17 June 2019**

**Conformance Standard**

ISO/IEC 17025:2017

General requirements for the competence of testing and calibration laboratories

**Laboratory Services Summary**

2.64

Asbestos

Operations Manager Authorisation:		Issue 12	Date: 14/02/24	Page 1 of 2
--------------------------------------	--	----------	----------------	-------------



Chemsafety Ltd  
Chemical Testing Laboratory  
**SCOPE OF ACCREDITATION**

Accreditation Number 1328

**2.64 Asbestos**

**(a) Identification**

Bulk identification of asbestos using polarised light microscopy

AS 4964:2004 Method for the Qualitative Identification of Asbestos in Bulk Samples

**(b) Fibre Counts**

Determination of airborne asbestos fibre counts on membrane filters

NOHSC:3003: National Occupational Health and Safety Commission Guidance Note on the Membrane Filter Method for Estimating Airborne Asbestos 2<sup>nd</sup> Edition [NOHSC:3003 (2005)]

**(c) Field Sites**

Establishment and operation of field sites for the purposes of asbestos fibre counting, in accordance with procedures listed in 2.64(b)

**(d) Field Testing**

Performance of field testing for the purposes of asbestos fibre counting, in accordance with procedures listed in 2.64(b)

Operations Manager Authorisation:		Issue 12	Date:14/02/24	Page 2 of 2
--------------------------------------	--	----------	---------------	-------------